



Eckert & Ziegler
Nuclitec GmbH

Gieselweg 1
38110 Braunschweig
Germany

Tel. +49 5307 932-0
Fax +49 5307 932-293

24 June 2011

CERTIFICATE

No. 145603 - TR 925

for a Sealed Radioactive Source

Source Type: Reference Source

Product code SIR1211
Drawing VZ-599-001
Source no. TR 925
Nuclide/s Strontium-90

Technical Data

Activity 43.9 kBq
Uncertainty* 6 %
Reference date 23 June 2011 at 12:00 UTC
Traceability* Defined on page 2
Radioactive impurities Related to Sr-90 (equal 100 %) the following radioactive impurities were detected: none

Leakage and Contamination Test/s

Test method/s* |
Test/s passed on 24 June 2011

Additional Information

ISO classification* C.11111
Remark ---

* see page 2 for explanation

Eckert & Ziegler
Nuclitec GmbH

i. A. V. A.
Production Manager

